



The External Quality Assessment Schemes in Clinical Immunology : EQAI  
**โครงการประเมินคุณภาพห้องปฏิบัติการภูมิคุ้มกันวิทยาโดยองค์การภายนอก**

Department of Clinical Microbiology and Applied Technology, Faculty of Medical Technology, Mahidol University, Bangkok, Thailand 10700  
 ภาควิชาชีววิทยาคลินิกและเทคโนโลยีประยุกต์ คณะเทคโนโลยีการแพทย์ มหาวิทยาลัยมหิดล เลขที่ 2 ถนนวิภาวดี แขวงศิริราช เขตบางกอกน้อย กรุงเทพมหานคร 10700  
 Tel. 080-982-5570 Fax. 0-2412-4110 โทรศัพท 080-982-5570 โทรสาร 0-2412-4110



**TENTATIVE DESIGNATED RESPONSE FOR SYPHILIS SEROLOGY PROGRAM**

**TRIAL 01-20-2**

Control material No.	01-20-2-1	01-20-2-2	01-20-2-3	01-20-2-4
Non treponemal test	NS	Reactive	Reactive	Non-Reactive
Treponemal test	Non-Reactive	Reactive	Reactive	Non-Reactive

**TENTATIVE DESIGNATED RESPONSE FOR HEPATITIS B VIRUS SEROLOGY PROGRAM**

**TRIAL 02-20-2**

Control material No.	02-20-2-1	02-20-2-2	02-20-2-3	02-20-2-4
HBsAg	Negative	Negative	Positive	Negative
Anti-HBs (all methods) mIU/ml	Positive 1528.63,772.43nIQR	Positive 0.92,1.48nIQR	Negative 11071.05,2567.66nIQR	Positive 11090.02,3314.35nIQR
Anti-HBc	Positive	Negative	Positive	Negative
HBeAg	Negative	Negative	Positive	Negative
Anti-HBe	Positive	Negative	Negative	Negative